

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 18 September 2015
to 17 September 2018
Issued on 18 December 2015



ISO/ IEC 17025
CL 002- 01

Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 002 -01

Metrology Division
Sri Lanka Standards Institution
No 17, Victoria Place
Elvitigala Mawatha, Colombo 08

Scope of Accreditation: Performing Mechanical calibration (Mass & Force) and Thermal calibration as per the calibration methods appearing in the schedule.

The Laboratory is accredited for the following calibrations.

Sl	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	CMC approximately at 95% confidence level	Location
Mechanical Calibration (Force)						
01	Calibration of Force-providing instruments used for the verification of uni axial testing machines	ISO 376:2011	8kN-100kN	Proving Ring- 0.1 division 0.2 Load Cells- 0.00001 mV/ V	2.0 x 10 ⁻¹	SLSI
02	Verification & Calibration of Force Measuring Systems	ISO 7500-1: 2004	1) 20kN – 2000kN (Compression) ii) 20 N/ 5 kN (Tension)	Proving Ring- 0.1 division Load Cells- 0.001 mV, 0.1 kN & 1 kgf	2.2 x 10 ⁻¹	SLSI

[Signature]
Page 01 of 02

SI	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	CMC approximately at 95% confidence level	Location
Mass						
02	Mass/ weight/ Weights (Class F1 & Below Class F1)	DM/M/TM/02 based on OIML R-111: 2004 (Double Substitution method ABBA)	1mg - 50 mg	Not applicable	3 µg	In-house
			100 mg		4 µg	
			200 mg		5 µg	
			500 mg		6 µg	
			1 g		7 µg	
			2 g		9 µg	
			5 g		11 µg	
			10 g		14 µg	
			20 g		18 µg	
			50 g		0.03 mg	
			100 g		0.04 mg	
			200 g		0.08 mg	
			500 g		0.20 mg	
			1 kg		0.39 mg	
			2 kg		2 mg	
			5 kg		6 mg	
10 kg	8 mg					
20 kg	15 mg					
	Mass/ weight/ Electronic Balance	DM/M/TM/03 Based on calibration of weights and balance published by National measurement laboratory, Australia	0 g-20 g	0.01 mg	0.08 mg	In-house / Site
			20 g- 200 g	0.01 mg	0.11 mg	
			200g-500g	0.1 mg	0.27 mg	
			500g-1kg	1.0 mg	2 mg	
			1kg-5kg	10 mg	20 mg	
			10kg-20kg	100 mg	41 mg	
			50kg-150kg	1.0 g	0.5 g	
			150kg-200kg	2.0 g	0.8 g	

(Handwritten signature)

SI	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	CMC approximately at 95% confidence level	Location
Temperature						
03	Calibration of Liquid- in-glass thermometers	DM/T/TM-01	-80 °C - 50 °C 50°C - 350 °C 350 °C - 550 °C	0.01 °C	0.05 °C 0.06 °C 0.08 °C	SLSI / In- situ
	Calibration of Dial Thermometers	DM/T/TM-02	-80°C - 550 °C	0.1 °C	0.2 °C	SLSI / In- situ
	Calibration of Digital thermometers with a sensor	DM/T/TM-03	-80 °C - 50 °C 50 °C - 350 °C 350°C - 500 °C 500°C-1000 °C 1000 °C -1200 °C	0.01 °C 0.1 °C	0.05 °C 0.06 °C 0.08 °C 0.8 °C 1.8 °C	SLSI / In- situ
	Evaluation of Performance of Autoclaves	DM/T/TM-04	50°C - 150 °C	0.1 °C	0.6 °C	SLSI / In- situ
	Evaluation of Performance of furnaces	DM/T/TM-05	200°C - 1000°C	1 °C	1 °C	SLSI / In- situ
	Evaluation of Performance of liquid baths	DM/T/TM-06	-30 °C -200 °C	0.01 °C	0.05 °C	SLSI/ In- situ
	Evaluation of Performance of ovens	DM/T/TM-07	30 °C - 200 °C	0.1 °C	0.8 °C	SLSI / In- situ
	Evaluation of Performance of Incubators	DM/T/TM-08	0 °C - 60 °C	0.1 °C	0.6 °C	SLSI / In-situ
	Evaluation of Performance of Cold rooms	DM/T/TM-09	-80 °C - 20 °C	0.1 °C	0.6 °C	Site
	Calibration Thermocouples	DM/T/TM-10	0 °C - 960°C 960 °C - 1000 °C 1000°C -1200°C	0.1 °C	0.5 °C 0.9 °C 1.8 °C	SLSI
	Calibration of PRT's by comparison method	DM/T/TM-11	-80 °C - 0 °C 0 °C - 660°C	0.01 °C	0.02 °C 0.03 °C	SLSI


 Director/CEO
 Sri Lanka Accreditation Board for Conformity Assessment
 Colombo



Valid from 10 February 2016
to 17 September 2018
Issued on 10 February 2016

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/ IEC 17025
CL 002- 01

Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 002 -01

Metrology Division
Sri Lanka Standards Institution
No 17, Victoria Place
Elvitigala Mawatha, Colombo 08

**Scope of Accreditation: Performing Mechanical calibration (Length and Pressure)
And Temperature mapping as per the calibration methods
appearing in the schedule.**

The Laboratory is accredited for the following calibrations.

Sl	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	CMC approximately as 95% confidence level	Location
Length						
01	Calibration of Digital External Micrometer	DM/L/TM/01	0 mm -25 mm	0.001 mm	± 0.001 mm	SLSI
02	Calibration of Mechanical External Micrometer	DM/L/TM/01	0 mm- 25 mm	0.001 mm	± 0.002 mm	SLSI
03	Calibration of Digital calipers	DM/L/TM/02	0 mm-150 mm	0.01 mm	± 0.01 mm	SLSI / In-situ
04	Vernier Caliper	DM/L/TM/02	0 mm-150 mm	0.02 mm	± 0.02 mm	SLSI / In-situ

SI	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	CMC approximately as 95% confidence level	Location
Pressure						
01	Calibration of compound gauges	DM/P/TM/01 (Based on DKD-R 6-1:2003)	-900mbar - 20 bar	0.1 bar	0.2 bar	SLSI / In-situ
02	Calibration of air pressure gauges	DM/P/TM/01 (Based on DKD-R 6-1:2003)	0 bar - 20 bar	0.1 bar	0.2 bar	SLSI / In-situ
03	Calibration of hydraulic pressure gauges	DM/P/TM/01 (Based on DKD-R 6-1:2003)	0 bar - 600 bar	0.1 bar	0.2 bar	SLSI / In-situ
Temperature Mapping						
	Temperature Mapping of Temperature controlled enclosures	DM/T/TM-12	-80°C-30°C (Volume = 20m ³) -80°C-30°C(20m ³ = volume = 100m ³)	0.1 °C	0.6 °C 0.7 °C	SLSI / In-situ



Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment
Colombo , Sri lanka