

AMENDMENT NO : 1 APPROVED ON 2008-05-26 TO SLS 516 PART 1 : 1991

**MICROBIOLOGICAL TEST METHODS
PART 1 – GENERAL GUIDANCE FOR ENUMERATION OF
MICROORGANISMS COLONY COUNT TECHNIQUE (FIRST REVISION)**

EXPLANATORY NOTE

This amendment is issued to incorporate incubation temperatures 22⁰ C and 37⁰ C other than 30⁰ C.

**AMENDMENT NO : 1 TO SLS 516 PART 1 : 1991
MICROBIOLOGICAL TEST METHODS
PART 1 – GENERAL GUIDANCE FOR ENUMERATION OF MICRO
ORGANISMS COLONY COUNT TECHNIQUE (FIRST REVISION)**

Title of the standard to be corrected as

Microbiological Test Methods
Part 1 General guidance for enumeration of microorganisms colony count technique.

Page 3

Clause 1 – Scope

Include temperatures 22 °C and 37 °C other than 30 °C

Clause 3 – Definitions

Include temperatures 22 °C and 37 °C other than 30 °C

Page 4

Clause 5.6

Delete and substitute with

Incubator, controlled at 22 ± 1 °C or 30 ± 1 °C or 37 ± 1 °C

AMD 371

Page 10

Clause 9.1.5

Include following

$22 \pm 1^{\circ}\text{C}$ for 72 ± 3 h

$37 \pm 1^{\circ}\text{C}$ for 48 ± 2 h

Clause 9.2.6

Include following

$22 \pm 1^{\circ}\text{C}$ for 72 ± 3 h

$37 \pm 1^{\circ}\text{C}$ for 48 ± 2 h

Page 12

Under example

Delete the first sentence and substitute with

A microorganism count at 22°C or 30°C or 37°C gave the following results.